DAIRY · MILK TESTING · LOG

Milk analyser cleaning log

Residual fat, protein and scale in a milk analyser’s measuring cell drift the fat and SNF readings that decide what a farmer is paid. Run Lacto Daily, a mild alkaline wash, at the end of every testing day, and Lacto Weekly, an acidic wash, once a week. Always follow the analyser maker’s cleaning instructions.

CENTRE	ANALYSER MAKE / MODEL	SERIAL NO.	MONTH

Cleaning schedule

When	What to do	Product
After every testing session	Run the analyser’s cleaning cycle with the daily cleaner	Lacto Daily
Once a week	Run the deep-clean cycle with the acidic cleaner, then the daily rinse	Lacto Weekly
After cleaning	Check a reference sample before testing resumes	None

Daily and weekly record

Date	Daily clean (Lacto Daily)	Weekly clean (Lacto Weekly)	Reference sample OK?	Done by

Date	Daily clean (Lacto Daily)	Weekly clean (Lacto Weekly)	Reference sample OK?	Done by

Sign-off

OPERATOR	SUPERVISOR	QA
Name, signature and date	Name, signature and date	Name, signature and date

Use the dose, temperature and contact time on each product’s technical data sheet (TDS) and your site’s validated cleaning programme. Read the SDS and wear the PPE it lists. Never mix chemicals. This sheet is a guide; follow your plant QA plan and current FSSAI requirements. SDS and TDS: www.innovachem.in/resources/sds-tds/